

DISSECTING FORCEPS NO TEETH 15 CM

Ref. IP26215



27/08/2024 03:25:13 - For information purposes only, can be subject to change

GSH | 30 rue Jean de Guiramand
13290 Aix-en-Provence | France

www.gsh-med.fr

+33(0)4 42 90 31 31

info@gsh-med.fr

DISSECTING FORCEPS NO TEETH 15 CM



Ref. IP26215

Used during surgical procedures or care. Convenient for removing tissues and extracting foreign bodies.

Description

Holtex instruments are all made of premium stainless steel (AISI 304, 410, 420 and 420A, depending on the type of instrument) to ensure good durability and corrosion resistance.

The steels come into contact with human tissues, so they must comply with international standards and regulations.

Holtex guarantees you medical devices that comply with Directive 93/42/EC and ISO 13485 (2016).

- In the "ATTACHED DOCUMENTS" tab, you can download the "Cleaning and sterilization" leaflet

Features

CE Marking	CE	Medical device	Class I
Bar code	3661040024028	HS Code	90189090
Warranty (months)	24	Country	Pakistan
DIMENSIONS AND WEIGHT			
Packed unit / Width (cm)	nc	Bare Unit / Width (cm)	0.7
Packed unit / Height (cm)	nc	Bare Unit / Height (cm)	0.3
Packed unit / Length (cm)	nc	Bare Unit / Length (cm)	15
Packed unit / Weight	0.033	Bare Unit / Weight (kg)	0.033
SPECIFIC STORAGE CONDITIONS			
Minimum temperature	NA	Max temperature	NA
Minimum humidity	NA	Maximum humidity	NA
OTHER FEATURES			
Packed unit / Quantity	1		

27/08/2024 03:25:13 - For information purposes only, can be subject to change

GSH | 30 rue Jean de Guiramand
13290 Aix-en-Provence | France

www.gsh-med.fr

+33(0)4 42 90 31 31

info@gsh-med.fr