

# **DIRECTOR AND TONGUE TIE 18 CM**

Ref. IS02418













30/07/2025 02:24:08 - For information purposes only, can be subject to change





## **DIRECTOR AND TONGUE TIE 18 CM**



#### Ref. IS02418

### Convenient for protecting organs during surgery

### **Description**

Holtex instruments are all made of premium stainless steel (AISI 304, 410, 420 and 420A, depending on the type of instrument) to ensure good durability and corrosion resistance.

The steels come into contact with human tissues, so they must comply with international standards and regulations.

Holtex guarantees you medical devices that comply with Directive 93/42/EC and ISO 13485 (2016).

- In the "ATTACHED DOCUMENTS" tab, you can download the "Cleaning and sterilization" leaflet

#### **Features**

CE Marking	CE	Medical device	Class I
Bar code	3661040030067	HS Code	90189084
Warranty (months)	24	Country	Pakistan
DIMENSIONS AND WEIGHT			
Packed unit / Width (cm)	nc	Bare Unit / Width (cm)	2
Packed unit / Height (cm)	nc	Bare Unit / Height (cm)	0.5
Packed unit / Length (cm)	nc	Bare Unit / Length (cm)	17.7
Packed unit / Weight	0.009	Bare Unit / Weight (kg)	0.009
SPECIFIC STORAGE CONDITIONS			
Minimum temperature	NA	Max temperature	NA
Minimum humidity	NA	Maximum humidity	NA
OTHER FEATURES			
Packed unit / Quantity	1		

30/07/2025 02:24:08 - For information purposes only, can be subject to change

GSH | 30 rue Jean de Guiramand 13290 Aix-en-Provence | France









www.gsh-med.fr