

ROCHESTER-PEAN FORCEPS STRAIGHT NO TEETH 15 CM

Ref. IP02215



CE

DISPOSITIF MEDICAL
CLASSE
I
MEDICAL DEVICE

GARANTIE
2 ANS
2
2 YEARS
GUARANTEE

STAINLESS
INOX
STEEL

07/04/2026 00:24:01 – For information purposes only, can be subject to change

ROCHESTER-PEAN FORCEPS STRAIGHT NO TEETH 15 CM



Ref. IP02215

Convenient to stop bleeding without damaging fragile tissues during an operation. Also used to perform care (dressings, etc.).

Description

Holtex instruments are all made of premium stainless steel (AISI 304, 410, 420 and 420A, depending on the type of instrument) to ensure good durability and corrosion resistance.

The steels come into contact with human tissues, so they must comply with international standards and regulations.

Holtex guarantees you medical devices that comply with Directive 93/42/EC and ISO 13485 (2016).

- In the "ATTACHED DOCUMENTS" tab, you can download the "Cleaning and sterilization" leaflet

Features

CE Marking	CE	Medical device	Class I
Bar code	3661040021560	HS Code	90189084
Warranty (months)	24	Country	Pakistan
DIMENSIONS AND WEIGHT			
Packed unit / Width (cm)	nc	Bare Unit / Width (cm)	7.5
Packed unit / Height (cm)	nc	Bare Unit / Height (cm)	0.7
Packed unit / Length (cm)	nc	Bare Unit / Length (cm)	15.6
Packed unit / Weight	0.044	Bare Unit / Weight (kg)	0.043
SPECIFIC STORAGE CONDITIONS			
Minimum temperature	NA	Max temperature	NA
Minimum humidity	NA	Maximum humidity	NA
OTHER FEATURES			
Packed unit / Quantity	1		

📅 07/04/2026 00:24:01 – For information purposes only, can be subject to change